Sample application

For Distributor Only



(Benzene, Toluene, Xylene, etc.)

Recycled products (Recycled oil, Waste oil, etc.)






	Amount, mg	n	Result, µg/mL or µg/g	RSD, %
Diesel	20 (μL)	3	52	2.1
Heavy oil	20 (μL)*1	3	2350	1.6
Lubricant oil	20 (μL)*1	3	375	1.8
Polycarbonate	13	5	2.5	4.5
Polyethylene	12	5	27	3.8
Epoxy resin	11	5	31	1.2
Tonner	8	5	355	1.5
Rubber	5	3	270	1.2
Pulp	3	5	3750	2.1

* Diluted by toluene

Sulfur(UVFL) sample application

	Amount, mg	n	Result, μg/mL or μg/g	RSD, %
Naphtha	10 (μL)	5	181	0.6
Diesel	10 (μL)	3	133	0.6
Kerosene	10 (μL)	3	25	1.2
Gasoline	10 (μL)	3	145	1.8
Lubricant oil	10 (μL)	3	2870	1.2
Heavy oil	10 (μL)	3	1340	0.5
PBT	30	5	303	2.6
Pulp	5	3	206	1.6

Chlorine(MCD) sample application



Sulfur(MCD) sample application

	Amount, mg	n	Result, μg/mL or μg/g	RSD, %
Lubricant oil A	5 (μL)	3	1.2 (%) 3.5	
Lubricant oil B	10 (μL)	3	0.76 (%)	3.5
Lubricant oil C	10 (μL)	3	520	4.3
Rubber	15	3	740	3.2
Epoxy resin	15	3	130	2.4
Coal	10	3	320	6.1
Coke	10	3	570	3.2
Crude oil	5	3	120	3.1

Note:

Follow instructions in manuals to correctly install, connect and operate the instruments. Contents of catalogues are subject to change without prior notice when improvements are made in performance. The actual color of the goods may appear deferent from color printed. All screen images are inset synthesis.

Company and product names contained herein are the trademarks or registered trademarks of the company concerned.



Read through the user's manual first before installing, piping, wiring, and operating the instrument, then always follow the manual for proper operation.

Nittoseiko Analytech

7-10-1 Chuo-rinkan, Yamato, Kanagawa, 242-0007, Japan TEL: +81-46-278-0056 FAX: +81-46-278-0057 URL: https://www.n-analytech.co.jp/global/CAT No.310101223051E

Pantech Business Hub Singapore 128383 Tel: 65-67790007 Fax: 63520877, 63520885 Email:sales@sintechscientific.com Website: www.sintechscientific.com

No 194 Pandan Loop #04-11

Sintech Scientific (S.E.A.) Pte Ltd





Nittoseiko Analytech Co., Ltd.



NSX-5000H

NSX-5000H series contributes to the reduction of environmental load through advanced technologies. It enables automatic, safe and highly sensitive measurement of trace nitrogen, sulfur and chlorine in liquid or solid samples. Can be used in a wide range of applications such as quality control of petroleum/ chemical/ recycled products or automotive fuels and environmental analysis.



User-friendly software

The improved user interface can be adjusted to the requirements of different operators.



Simple mode allows easy operation in a routine environment, using predefined methods and guided start-up and shut-down procedures.

Automatic start-up and shut-down

Instrument start-up and shut-down procedures can be automated including flexible temperature and gas flow settings.

-	-	-	200													
				-	568 applic. note LPG	KÚ/S	561,	Myl organ	c(3) ×							_
2	O	P	тераге		Stant read	Callo	ratio	in curve read	Parame	art .	Step Editor					
	4		Main Typ		Sub type			ert ,	Sample ID		Calibration ID	B No.	Volume		Injector	7
n	1		Standby	_	Electric Furnace On		L	0.0	- Francisco		22111	-	250000		100000	
	2		Standby	T	Gas On	-	L	_		L		_				
9	3		Diank.	7	Conditioning	_	5	20		E		1	40 pt.	×	ASC-550L	4
	*		Blank	3	Conditioning		3	- 1		E		-1	40 pL	Ĕ	ASC-550L	4
	3	-	Standard	-	Calibration		8	- 24			5.88 ppm S LFV		150 pt.	×	QI-560	4
	•	_	Standard	-	Calibration	_	8	- 2			5.80 ppm S LF v		180 pt.	Y	GI-560	4
1	2	_	Standard	7	Calibration	_	8	-			5.88 ppm S LFIII		240 pt.	4	C1-660	4
1			Standard	×	Calibration		3			466	5.88 ppm S.LF	_	240 pt	×	GI-560	4
	,		Standard	*	Calibration		5	1			5.88 ppm 5 LFV		270 pl.	4	QI-560	4
	10	_	Standard	_	Calibration		\$	- 1		12	5.80 ppm S LF!=		270 pt.	9	GI-560	1
	11	0	Standby	14	Electric Furnace Of	9										
	12	0	Standby		Dry Gas Of	- Ev										
	13		Standby		Xenon Lamp Of	1	Г			Г						
	14	0	Standby		Gas Of	- 24	Г			Е						

Import/Export function

Measurement related data and software related settings can be exported and (re-)imported. This is not only for data security and backups, but makes transferring methods and user settings to other computers and instruments easy.



Link to LIMS

Automatic output of measurement data to the network allow seamless integration into laboratory data management.

Advanced mode

Advanced mode gives full flexibility for method development, customization and access to quality control and auditrial functions.

Audit trail

All operations and instrument conditions are recorded and allow full traceability of the measurement process.

ile E	est V	iew Control Run Tool Help			
-		terr control non-room resp			
per	2	MOOE I	0		
-	1.1	THE REAL PROPERTY.			
-8	Sct.	Start Calibration curve Blank	Auditrail		
遪			1 4 4	Tak .	Tree Cont
49	No.	Type Date 16 Into 4/22/2022 3 39 04 PM	Sorting	Message Step deletion (20220376 17_001_GFS60 applic no	UseriD
2	20	R Info. 4/22/2022 3:39:04 PM	Step	Step deletion (20220316 17 001 GI-560 applie no	Administrati
	21	Info 4/22/2022 3 39 04 PM	Step	Step deletion (20220318 17_001_QI-580 applic no	Administrat
	22	Info 4/22/2022 3:39:04 PM	Step	Step deletion (20220316 17_001_GI-560 applic no	Administrat
AD:	23	Info 4/22/2022 3:39/04 PM	Step	Step detetion (20220316 17 001 GI-560 applic no	Administrat
	24	Info 4/22/2022 3 39 24 PM	Step	Step deletion (20220316 17 001 GI-560 applic no	Administrat
-	25	Info 4/22/2022 3/38/24 PM	Step	Step detetion (20220316 17 001 GI-560 applic no	Administrat
Q	26	Info 4/22/2022 3:40:17 PM	Step	Step addition (20220316 17 001 Gi-560 applic no	Administrat
	27	Info 4/22/2022 3:40:24 PM	Step	Step edit (20220316 17 001 GI-560 applic note L	Administrat
		Info 4/22/2022 3:40:24 PM	Step	Step edit (20220316 17, 001_GI-560 applic note L	A CONTRACTOR OF THE PARTY OF TH
	28	Info 4/22/2022 3:40:24 PM	Step	Step edit (20220316 17_001_GI-560 applic. note L	Administrat Administrat
	30	Info 4/22/2022 3:40 35 PM	Step	Step edit (20220316 17_001_GI-560 applic note L	Administrat
	31	Info 4/22/2022 3:40:35 PM	Sten	Step edit (20220316 17_001_GI-860 applic note L	Administrat
	32	Info 4/22/2022 3:40:47 PM	Step	Step addition (20220316 17_001_GI-560 applic.no	Administrat
	33	Info 4/22/2022 3:40:50 PM	Step	Step edit (20220316 17 001 GF560 applic note L	Administrat
	34	Info 4/22/2022 3:40:50 PM	Step	Step edit (20220316 17_001_GF560 applic. note L	Administrat
		The state of the s		Step edit (20220316 17_001_GI-560 applic note L	
	35 36	Info 4/22/2022 3:40:50 PM Info 4/22/2022 3:41:06 PM	Step	Step 4dit (20220316 17_001_GI-560 applic. note L	Administrati Administrati

User management

Multiple users with different access levels help to protect data from unintended changes and support a cleanand easy to operate user interface.

Edit User Account					
User ID	User ID cannot	contain to	e following characters, 17:*1	5743-1	
Password:	_				
Password (Confirm)					
Password hint:					
Level:	Analytier	4	Schedules can be prepared achedules can not be delet		
User Information					
UserName:			Phone:		
Company:			Fax:		
Department:			Exension:		
Office:			Monte:		
Job Title			MalAggress		
Comment:					
Authority setting					
Export	Parameter s	fain Menu	Explorer Schedule Sim	rple Mode Advanced	Mode Others
Import	VF-500 MCI	D-500 S	D-500 ND-500 ASC-550L	PC	
Return to default		ater Tem rater Tem rater Tem	p. Outlet (°C) p. Intet (Standby) (°C) p. Outlet (Standby) (°C)		û
☐This user ID is displ	ayed in the logon	list		OK	Cancel

Easy Daily maintenance

Unique openable furnace and user-friendly connectors make routine visual checks and maintenance of the pyrolysis tube a matter of seconds, thus increasing consumable lifetime and performance.

01



Low running cost

Less gas consumption than before. It is also available to shut off the gas at the end of measurement automatically.


Rapid analysis with automatic boat control

Putting the sample boat into the electric furnace with an appropriately controlled speed prevents incomplete combustion and enables rapid measurement. New programs can be created easily.

Reanalysis function

*Available at SD-500 or ND-500

Manual peak reintegration is possible after measurement to correct and optimize results as required.



Vial tray for liquid samples

Electronic cooling (Peltier)

viscous samples

30 ma or less

Boat tray for solid and non-volatile or

10 to 35 °C (No fluctuation during measurement)

500 (W) \times 460 (D) \times 600 (H) mm, Approx. 27kg

99 positions + end position + cooling position

AC 100/115/230/240 V, 50/60 Hz, 192 VA

Options

Automatic Sample Changer for Liquid and Solid ASC-570LS



Samples tray type

Ambient Temperature

Sample boat quantity Boat operation

Boat cooling

Power supply

Volume

Mass and weight

		Sample boat material	Ceramic (Standard)		
	ъ	Sample container	Vial with a septum, 4 mL or 2 mL capacity		
	Liquid	Sample vial quantity	4 mL- 84 pcs, 2 mL- 120 pcs		
	_	Injection system	Gastight Microsyringe 25, 50, 100 µL		
	Sa	ample type	Solid/Liquid		
		ample type	Solid/Liquid Solid: 30 mg or less, Liquid: 100 µL or less		
	Sa		'		
	Sa	ample volume	Solid: 30 mg or less, Liquid: 100 μL or less		

Automatic Boat Controller ABC-500

	Sample type	Solid/Liquid		
N. A. M.	Sample volume	Solid: 30 mg or less, Liquid: 100 µL or less		
	Sample boat material	Ceramic, Quartz		
	Boat cooling	Electronic cooling (Peltier)		
	Power supply	AC 100/115/230/240 V, 50/60 Hz, 40 VA		
	Mass and weight	450 (W) × 250 (D) × 180 (H) mm, Approx. 9 kg		



Trace Nitrogen Analyzer NSX-5000H /ND



Trace Sulfur Analyzer NSX-5000H /SD





Standard test method

ASTM D4629, D5762, D6069, D7184, JIS K2609



Measurement principle


Sample is loaded by sample boat into a high-temperature (900 to 1000 °C) pyrolysis tube with argon carrier gas. After nitrogen compounds in the sample are pyrolyzed, it is combusted, oxidized, and converted to nitric oxide (NO). After removing moisture from the combustion gas by a dehumidifier (tube dryer), the following oxidation reaction occurs by the reaction of NO with ozone.

$NO + O_3 \rightarrow NO_2 + O_2 + hv$

By this reaction, 590 to 2,500 nm of emitting light is generated. The optical intensity of this light is proportional to the NO concentration at a wide wave-length range. After emitted light is detected by a photomultiplier tube and signal processing is performed, an area value is obtained. Using the calibration curve, the total nitrogen concentration in the sample is calculated. CO₂ and SOx generated during oxidative combustion can cause interferences. However, the interferencecan be decreased to an irrelevant level by applying reduced pressure inside the reaction chamber.

○ NO-500 □ NF-500 □ NO-500

CLD Nitrogen Diagram



Specification for Trace Nitrogen Analyzer NSX-5000H /ND

	Nitrogen measurement				
Sample	Liquid (non-aqueous)/ Solid				
Analytical method	Oxidative combustion and chemiluminescence detection				
Furnace	Max. 1,100 °C, Openable electric furnace, 2 heating zones				
Measuring range*	Liquid: 0.05 ~ 10,000 μg/mL, Solid: 0.5 ~ 10,000 μg/g				
Sample amount	Liquid: Max 100 μL, Solid: Max 30 mg				
Measurement time	Approx. 8 min * It depends on the sample volume and measurement condition.				
Vacuum pump	Diaphragm type dry vacuum pump				
Gas supply	Argon: Purity 99.98 % or more, 0.3 ± 0.1 MPa, Oxygen: Purity 99.7 % or more, 0.3 ± 0.1 MPa				
Power supply	HF-500: AC100/115 V (50/60 Hz), 1100 VA, AC230/240 V (50/60 Hz), 1800 VA ND-500: AC100~240 V, 300 VA				
Mass and weight	HF-500: Approx. 320(W) ×430(D) ×500(H) mm, Approx. 25 kg ND-500: Approx. 220(W) ×375(D) ×500(H) mm, Approx. 22 kg				

^{*}Actual measuring range depends on the sample, the matrix, the purity of reagents, gases and the condition of the unit.

Standard test method

ASTM D5453, D7183, ISO 20846, JIS K2541-6



Measurement principle


Sample is loaded by sample boat into a heated (800 to 1100 °C) pyrolysis tube with argon carrier gas. After sulfur compounds in the sample are pyrolyzed, they are oxidized by O₂ gas. The produced SO₂ gas is excited by irradiating ultraviolet light v1 (190 to 230 nm). On returning to ground state, this SO₂* emits energy as fluorescent ray.

$$SO_2 + hv_1 \rightarrow SO_2^* \rightarrow SO_2 + hv_2$$

This fluorescent ultraviolet light v2 (300 to 400 nm) is received by a photomultiplier tube and the signal is integrated. The sulfur concentration is obtained using a calibration curve preliminarily drawn with standard solutions.



UVFL Sulfur Diagram



Trace Sulfur Analyzer NSX-5000H /SD (/ND)

	Sulfur measurement	Nitrogen measurement (optional)					
Sample	Liquid (non-aqueous) / Solid						
Analytical method	Oxidative combustion and UV fluorescence detection Oxidative combustion and chemiluminesc						
Furnace	Max. 1,100 °C, Openable electric furnace, 2 heating zones						
Measuring range*	Liquid: 0.05 ~ 10,000 μg/mL Liquid: 0.05 ~ 10,000 μg/mL Solid: 0.1 ~ 10,000 μg/g Solid: 0.1 ~ 10,000 μg/g						
Sample volume	Liquid: Max 100 μL, Solid: Max 30 mg						
Measurement time	Approx. 8 min						
Vacuum pump		Diaphragm type dry vacuum pump					
Gas supply	Argon: Purity 99.98 % or more, 0.3 ± 0.1 MPa, Oxygen: Purity 99.7 % or more, 0.3 ± 0.1 MPa						
Power supply	HF-500: AC100/115 V (50/60 Hz), 1100 VA, AC230/240 V (50/	60 Hz), 1800 VA, SD/ND-500: AC100-240 V, 300 VA					
Mass and weigtht	HF-500: Approx. 320(W) ×430(D) ×500(H) mm, Approx. 25 kg SD-500: Approx. 220(W) ×375(D) ×500(H) mm, Approx. 21 kg , ND-500: Approx. 220(W) ×375(D) ×500(H) mm, Approx. 22 kg						

^{*}Actual measuring range depends on the sample, the matrix, the purity of reagents, gases and the condition of the unit.

03

^{*}Sulfur and Nitrogen simultaneous measurement is available by connecting SD-500 in series with ND-500.

Trace Chlorine/Sulfur Analyzer NSX-5000H /MCD





Chlorine: ASTM D4929, D5808, D7457

Sulfur: ASTM D3120, D3961

Measurement principle

Chlorine

Samples are combusted in an argon/oxygen atmosphere. The resulting hydrogen chloride is transferred into a titration cell where it is automatically titrated by silver ions generated by electrolysis. The amount of chlorine is calculated from the quantity of electrical charge required for the titration.

$HCI + Ag^+ \rightarrow AgCI \downarrow + H^+$ $Ag \rightarrow Ag^+ + e^-$


*Chlorine measurement in inorganic samples requires optional pyrolysis tube and humidification kit.

Sulfur

Samples are combusted in an argon/oxygen atmosphere. The resulting sulfur dioxide is transferred into a titration cell where it is automatically titrated by triiodide ions generated by electrolysis. The amount of sulfur is calculated from the quantity of electrical charge required for the titration.

 $SO_2 + I_3 + H_2O \rightarrow SO_3 + 3I^- + 2H^+$ $3I^- \rightarrow I_3 + 2e^-$

MCD Diagram for Chlorine (/Sulfur)



ABC 500 O O O O O

Specification for Trace Chlorine/Sulfur Analyzer NSX-5000H/ MCD

	Chlorine measurement	Sulfur measurement		
Sample	Liquid	I / Solid		
Analytical method	Oxidative combustion and microcoulometric detection			
Furnace	Max. 1,100 °C, Openable elec	ctric furnace, 2 heating zones		
Detection electrode	Silver electrode	Platinum electrode		
Measuring range*	Liquid: 0.1 ~ 5,000 μg/mL Solid: 0.5 ~ 50,000 μg/g	Liquid: 0.2 ~ 1,000 μg/mL Solid: 1.0 ~ 5,000 μg/g		
Sample size	Liquid: Max. 100 μL	uL, Solid: Max. 30 mg		
Measurment time	Approx. 10minut	tes for 2 µg Cl/S		
Gas supply	Argon: Purity 99.98 % or more, 0.3 ± 0.1 MPa	Oxygen: Purity 99.7 % or more, 0.3 ± 0.1 MPa		
Power supply	HF-500: AC100/115 V (50/60 Hz), 1100 VA, AC230/240V (50/60 Hz), 1800 VA MCD-500: AC100-240V, 150 VA			
Mass and weight	HF-500: Approx. 320(W) ×430(D) ×500(H) mm, Approx. 25 kg	kg, MCD-500: Approx. 220(W) ×375(D)×500(H) mm, Approx. 14 k		

05

Adsorbable Organic Halogens (AOX) Analysis

By using activated carbon technology and coulometric titration technology, adsorbable organic halogens in environmental water such as river water, lake water, and various industrial wastewater can be measured easily, quickly, and with high accuracy.

Relevant standard test method

AOX: ISO 9652 TOX: EPA 9020

Measurement principle

1.Adsorption and rinse

A liquid sample is passed through two columns filled with activated carbon. AOX, adsorbable organic halogen, remains in the column after rinsing to remove inorganic halogen compounds.

2.Combustion

Activated carbon is extruded from the column into the boat, then, the boat is carried into the horizontal furnace. AOX is pyrolyzed in the furnace to generate hydrogen halide.

3.Titration

Hydrogen halide is titrated with Ag ions produced by electrolysis. The amount of halogen is calculated from the amount of electricity required for titration.

 $HX + Ag^+ \rightarrow H^+ + AgX \downarrow$ $Ag \rightarrow Ag^+ + e^-$

Total Organic Halogen Adsorption Unit Model TXA-04

A sample can be passed through a column filled with activated carbon, adsorbed, and rinsed with an aqueous nitrate solution. Adsorption and rinse are switched manually.

		100 mg - 100		
Injection	Disposable syringe 20 mL	3 3 3 3 3		
Flow method	Automatic syringe system	200		
Maximum flow volume	Samples: 300 mL (10 × n mL (n=1 to 30)) Rinse: 50 mL (5 × n mL (n=1 to 10))			
Flow rate	3.0 mL/min (by official methods)			
Syringe quantity	5 pcs (for flowing samples or washing in	norganic halide)		
Column	3mm ×40 mm two-stage glass column			
Filling volume of activated carbon	About 40 mg/column			
Power supply	AC100/115 V, 50/60Hz 50 VA AC 230/240 V, 50/60Hz 50 VA			
Mass and weight	330 (W) × 220 (D) × 500 (H) mm, Approx. 8 kg			


		Volume, mL	n	Result, µg/L	RSD, %
Dis	River water	50	3	78	4
	Distilled water	100	3	3.8	10
	Industrial waste water	20*	3	54 (mg/L)	0.8


^{*}Diluted by pure water

Material of sample boat and its characteristics

Appearance	Material	Description			
	Ceramics	Mainly used for solid sample measurement. It is weaker than quartz against heat changes, and repeated use may damage the port. On the other hand, it is characterized by its high strength against substances that corrode quartz, and the capacity of the sample that can be loaded from the porcelain port is large. It can be used for automatic measurement in solid measurement mode of ASC-570LS.			
	Quartz	It is mainly used for liquid sample measurement with quartz wool stuffed. Resistant to heat changes and suitable for repeated use. On the other hand, if a sample containing alkali metals or alkaline earth metals is measured, it will devitrify, so care must be taken when using it. Cannot be used for automatic measurement in solid measurement mode of ASC-570LS.			
	Sintered quartz	It has the same shape as the ceramic port and is made of quartz. Resistant to heat changes and suitable for repeated use. On the other hand, if you measure a sample containing alkali metals or alkaline earth metals, it will devitrify, so be careful when using it. It can also be used for automatic measurements in the solid measurement mode of the ASC-570LS.			
	Porcelain	Mainly used for solid sample measurement. Since it is relatively inexpensive, it is recommended to use it when the residue after burning is disposable. It cannot be used for automatic measurement in solid measurement mode of ASC-570LS.			
	Nickel (inner boat)	An inner port that is installed inside a ceramic or sintered quartz port to prevent corrosion of the port. Since corrosive components remain in the inner port, the inner port is disposable. For sulfur determination, we do not recommend using the inner boat as it may result in lower readings.			

06





In combination with ABC-500

^{*}Actual measuring range depends on the sample, the matrix, the purity of reagents, gases and the condition of the unit.